

Summary of Modifications/Changes in this Update

This Summary of Changes is for information only.
It is not a part of the referenced document, and should not be used for project documentation.

U.S. Department of Veterans Affairs ♦ Office of Construction & Facilities Management

DATE OF THIS VERSION:

January 1, 2017

TITLE OF DOCUMENT:

Ceiling Mounted Patient Lift System, 11 73 00

DATE OF VERSION BEING SUPERSEDED (old):

September 1, 2014

DESCRIPTION OF DOCUMENT (previous title, number, other identifying data):

Ceiling Mounted Patient Lift System, 11 73 00

SUMMARY OF CHANGES IN THIS VERSION:

1. Section 1.3. Quality Assurance: C. Certification of compliance with VA requirements shall be provided by an independent third party, Inspector of Record (IOR).
2. Section 1.4. Submittals: A Shop Drawings, Product Data, and Samples:
 - a. Shop Drawings shall show structural supports for ceiling mounted lift and provide structural calculations for attachments to structure.
 - b. A set of stamped and signed drawings shall be provided by a registered structural engineer and shall meet all code requirements for the jurisdiction having authority.
 - c. Shop drawings shall show obstructions and coordinate their relocation.
 - d. Manufacturer shall provide BIM for clash detection at the request of the VA or General Contractor.
3. Section 1.5 Applicable Publications:
 - a. D. International Electromagnetic Commission: has been changed to 60601-1-2 (2015) as of April 1st 2017.
 - b. E. VA Patient Safety Alert AL14-07: a editable PDF checklist is now provided on VHA's Patient Safety Website.

4. Section 2.1 Ceiling Track System:
 - a. As per ISO 10535 standards, installed rail shall be tested for 1.5 times greater than the motor's weight capacity.
5. Section 2.2 Lifting Unit:
 - a. B. Lifting Capacity shall be 550 lbs for non-bariatric lifts, and 1,000 lbs for bariatric lifts.
6. Section 2.4 Batteries:
 - a. B. for non-bariatric lifts a minimum of 40 transfers. For Bariatric lifts a minimum of 50 transfers.
7. Section 2.6 Straps and Sling:
 - a. Straps and Slings shall meet ISO 10535, Appendix A Guidelines.
8. Section 3.4 Inspection:
 - a. Periodic Inspection shall be provided by manufacturer on a yearly basis in compliance with ISO 10535.

1.